

DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g

Batch#: DFNY-MAUWOW-010926, Batch Size: 892 Units, LOT ID:

1A41203000001FC000000374

Date Sampled: 01/16/2026 16:44

Sample Received: 01/16/2026 17:52

Report Created: 01/29/2026 11:38

Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000372, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Results

394.64 mg/unit
D9-THC

356.43 mg/unit
THCa

ND
Total CBD

707.23 mg/unit
Total THC


773.65 mg/unit
Total Cannabinoids

2.159%
Total Terpenes

Tests Summary

Cannabinoids Tested	Moisture Pass	Microbials Pass
Water Activity Pass	Foreign Matter Not Tested	Homogeneity Not Tested
Terpenes Pass	Residual Solvents Pass	Mycotoxins Pass
Heavy Metals Pass	Pesticides Pass	

Dominant Terpenes



DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580
1050 Lower South St
New York, 10567
lpalazzo@jeeter.com
4424037621
OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g
Batch#: DFNJ-MAUWOW-010926, Batch Size: 892 Units, LOT ID:
1A41203000001FC000000374
Date Sampled: 01/16/2026 16:44
Sample Received: 01/16/2026 17:52
Report Created: 01/29/2026 11:38
Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000372, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Cannabinoids

SOP: SOP 801-NY

Instrument: Agilent HPLC 2518 -
2518 / Agilent Infinity II - 2779

Date/Time Tested: 01/19/2026
18:51

Tested

Analyte	LOQ (ug/mL)	%	mg/g	mg/unit
CBDV	1,537.53	ND	ND	ND
CBDa	1,537.53	ND	ND	ND
CBGa	1,537.53	0.22	2.24	4.49
CBG	1,537.53	0.44	4.38	8.76
CBD	1,092.45	< LOQ	< LOQ	< LOQ
THCV	1,092.45	0.14	1.44	2.88
CBN	1,227.20	0.16	1.58	3.16
D9-THC	1,227.20	19.73	197.32	394.64
D8-THC	1,227.20	ND	ND	ND
(6aR,9S)-d10-THC	1,092.45	ND	ND	ND
(6aR,9R)-d10-THC	1,227.20	ND	ND	ND
CBC	1,227.20	0.16	1.64	3.29
THCa	1,092.45	17.82	178.21	356.43
Total CBD		ND	ND	ND
Total THC		35.36	353.61	707.23
Total Cannabinoids		38.68	386.82	773.65

Report notes:

Total THC = THCa * 0.877 + Δ8-THC + Δ9-THC + (6aR,9S)-d10-THC + (6aR,9R)-d10-THC
Total CBD = CBDa * 0.877 + CBDT
Total Cannabinoids = Sum of all cannabinoids
LOQ = Limit of Quantitation; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory. OCMPPCL-2022-00001. Cannabinoid potency values for flower type products are reported by percentage of dry weight determined via loss on drying; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory. All results were generated by ISO certified methods to full state testing requirements. ND = Not Detected; NT = Not Tested; NR = Not Reported

Moisture

SOP: SOP 205-NY

Instrument: Shimadzu Moisture
Meter - 2560

Date/Time Tested: 01/19/2026
14:06

Pass

Analyte	Limit (%)	%	Status
Moisture	15	6.2	Passed

Microbials

SOP: SOP 401-NY SOP 418-NY

Instrument: 35C Incubator - 1032
/ 28C Incubator - 1034 / BioRad CFX
qPCR - 3195

Date/Time Tested: 01/18/2026
18:02

Pass

Analyte	LOQ (CFU/g)	Limit (CFU/g)	CFU/g	Status
Aerobic Bacteria	100		ND	Tested
E. Coli		0	ND	Passed
Yeast & Mold	100		ND	Tested
Salmonella		0	ND	Passed
Aspergillus Flavus		0	ND	Passed
Aspergillus Fumigatus		0	ND	Passed
Aspergillus Niger		0	ND	Passed
Aspergillus Terreus		0	ND	Passed

Report notes:

Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

Water Activity

SOP: SOP 206-NY

Date/Time Tested: 01/19/2026
14:06

Pass

Analyte	Limit (aw)	aw	Status
Water Activity	0.65	0.31	Passed

DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g

Batch#: DFNY-MAUWOW-010926, Batch Size: 892 Units, LOT ID:

1A41203000001FC000000374

Date Sampled: 01/16/2026 16:44

Sample Received: 01/16/2026 17:52

Report Created: 01/29/2026 11:38

Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000372, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Terpenes

SOP: SOP 620-NY

Instrument: Agilent GC-FID 1950 - 1950

Date/Time Tested: 01/19/2026 17:46

Pass

Analyte	LOQ (ug/mL)	Limit (ug/mL)	%	Status	Analyte	LOQ (ug/mL)	Limit (ug/mL)	%	Status
Total Terpenes		115,000	2.1593	Passed	Borneol	86		0.0098	Tested
Limonene	86		0.8418	Tested	Camphene	86		ND	Tested
b-Caryophyllene	86		0.3639	Tested	Sabinene	86		< LOQ	Tested
Linalool	86		0.2090	Tested	a-Phellandrene	86		ND	Tested
a-Humulene	86		0.1387	Tested	d-3-Carene	86		ND	Tested
a-Bisabolol	86		0.1192	Tested	a-Terpinene	86		ND	Tested
b-Myrcene	86		0.0830	Tested	p-Cymene	86		ND	Tested
Terpineol	86		0.0591	Tested	Eucalyptol	86		ND	Tested
trans-b-Farnesene	86		0.0514	Tested	Sabinene Hydrate	86		ND	Tested
Fenchol	86		0.0416	Tested	Fenchone	86		ND	Tested
b-Pinene	86		0.0392	Tested	Isopulegol	86		ND	Tested
Caryophyllene Oxide	86		0.0385	Tested	Camphor	86		ND	Tested
Geraniol	86		0.0335	Tested	Isoborneol	86		ND	Tested
g-Terpinene	86		0.0286	Tested	DL-Menthol	86		ND	Tested
cis-Nerolidol	86		0.0238	Tested	Nerol	86		ND	Tested
a-Pinene	86		0.0221	Tested	Geranyl Acetate	86		ND	Tested
Terpinolene	86		0.0151	Tested	a-Cedrene	86		ND	Tested
trans-Nerolidol	86		0.0150	Tested	Valencene	86		ND	Tested
trans-b-Ocimene	86		0.0144	Tested	Guaiol	86		< LOQ	Tested
Pulegone	86		0.0117	Tested	Cedrol	86		ND	Tested

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580
1050 Lower South St
New York, 10567
lpalazzo@jeeter.com
4424037621
OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g
Batch#: DFNY-MAUWOW-010926, Batch Size: 892 Units, LOT ID:
1A41203000001FC000000374
Date Sampled: 01/16/2026 16:44
Sample Received: 01/16/2026 17:52
Report Created: 01/29/2026 11:38
Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000372, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Residual Solvents

SOP: SOP 612-NY

Date/Time Tested: 01/27/2026 10:30

Pass

Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
1,2-Dichloroethane	3	5	ND	Passed
Acetone	2,500	5,000	< LOQ	Passed
Acetonitrile	205	410	ND	Passed
Benzene	1	2	ND	Passed
Butane	2,500	5,000	ND	Passed
Chloroform	30	60	ND	Passed
Ethanol	2,500	5,000	< LOQ	Passed
Ethyl Acetate	2,500	5,000	ND	Passed
Ethyl Ether	2,500	5,000	ND	Passed
DMSO	2,500	5,000	ND	Passed
Heptane	2,500	5,000	ND	Passed
Hexanes	145	290	ND	Passed
Isopropyl Alcohol	2,500	5,000	< LOQ	Passed
Methanol	1,500	3,000	< LOQ	Passed
Methylene Chloride	300	600	ND	Passed
Pentanes	2,500	5,000	ND	Passed
Propane	2,500	5,000	ND	Passed
Toluene	445	890	ND	Passed
Trichloroethane	750	1,500	ND	Passed
Xylenes	1,085	2,170	ND	Passed
1,1,1,2-Tetrafluoroethane (HFC-134a)	500	1,000	ND	Passed

Report notes:

LOQ = Limit of Quantitation; The reported result is based on a sample weight with the applicable moisture content for that sample; Unless otherwise stated all quality control samples performed within specifications established by the Laboratory. All results were generated by ISO certified methods to full state testing requirements. If DMSO and 1,1,1-Trichloroethane are reported, they are tentatively identified, but not quantitatively confirmed. ND = Not Detected; NT = Not Tested; NR = Not Reported.

DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580
1050 Lower South St
New York, 10567
lpalazzo@jeeter.com
4424037621
OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g
Batch#: DFNY-MAUWOW-010926, Batch Size: 892 Units, LOT ID:
1A41203000001FC000000374
Date Sampled: 01/16/2026 16:44
Sample Received: 01/16/2026 17:52
Report Created: 01/29/2026 11:38
Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000372, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Mycotoxins

SOP: SOP 808-NY

Instrument: Agilent LCMS 2546 - 2546

Date/Time Tested: 01/19/2026 16:41

Pass

Analyte	LOQ (ng/g)	Limit (ng/g)	ng/g	Status
B1	4.8		ND	Tested
B2	4.8		ND	Tested
G1	4.8		ND	Tested
G2	4.8		ND	Tested
Ochratoxin A	4.8	20.0	ND	Passed
Total Aflatoxins		20.0	ND	Passed
Total Mycotoxins			ND	Tested

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

Heavy Metals

SOP: SOP 250-NY

Instrument: Agilent 7800 ICP-MS - 2594

Date/Time Tested: 01/19/2026 11:46

Pass

Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
Antimony	0.188	2.000	ND	Passed
Arsenic	0.188	0.200	ND	Passed
Cadmium	0.188	0.200	ND	Passed
Chromium	0.188	110.000	< LOQ	Passed
Copper	0.225	30.000	6.428	Passed
Mercury	0.045	0.100	ND	Passed
Nickel	0.225	5.000	ND	Passed
Lead	0.188	0.500	ND	Passed

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. ND = Not Detected; NT = Not Tested; NR = Not Reported

DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580

1050 Lower South St

New York, 10567

lpalazzo@jeeter.com

4424037621

OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g

Batch#: DFNY-MAUWOW-010926, Batch Size: 892 Units, LOT ID:

1A41203000001FC000000374

Date Sampled: 01/16/2026 16:44

Sample Received: 01/16/2026 17:52

Report Created: 01/29/2026 11:38

Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000374, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Pesticides

SOP: SOP 814-NY

Instrument: Agilent LC-MS-MS 2755 - 2755 / Agilent GCMS 2541 - 2541

Date/Time Tested: 01/17/2026 18:13

Pass

Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
Abamectin	0.37	0.50	ND	Passed
Acephate	0.10	0.40	ND	Passed
Acequinocyl	0.10	2.00	ND	Passed
Acetamiprid	0.10	0.20	ND	Passed
Aldicarb	0.10	0.40	ND	Passed
Azoxystrobin	0.10	0.20	ND	Passed
Bifenazate	0.10	0.20	ND	Passed
Bifenthrin	0.10	0.20	< LOQ	Passed
Boscalid	0.10	0.40	ND	Passed
Carbaryl	0.10	0.20	ND	Passed
Carbofuran	0.10	0.20	ND	Passed
Chlorantraniliprole	0.10	0.20	ND	Passed
Chlorpyrifos	0.10	0.20	ND	Passed
Clofentezine	0.10	0.20	ND	Passed
Coumaphos	0.10	1.00	ND	Passed
Cyfluthrin	0.48	1.00	ND	Passed
Cypermethrin	0.10	1.00	ND	Passed
Daminozide	0.10	1.00	ND	Passed
Diazinon	0.10	0.20	ND	Passed
Dichlorvos	0.10	1.00	ND	Passed
Dimethoate	0.10	0.20	ND	Passed
Dimethomorph	0.10	1.00	ND	Passed
Ethoprophos	0.10	0.20	ND	Passed
Etofenprox	0.10	0.40	ND	Passed
Etoxazole	0.10	0.20	ND	Passed
Fenhexamid	0.10	1.00	ND	Passed
Fenoxycarb	0.10	0.20	ND	Passed
Fenpyroximate	0.10	0.40	ND	Passed
Fipronil	0.10	0.40	ND	Passed
Flonicamid	0.10	1.00	ND	Passed
Fludioxonil	0.10	0.40	ND	Passed
Hexythiazox	0.10	1.00	ND	Passed
Imazalil	0.10	0.20	ND	Passed
Imidacloprid	0.10	0.40	ND	Passed
Indole-3 Butyric Acid	0.12	1.00	ND	Passed
Kresoxim Methyl	0.10	0.40	ND	Passed
Malathion	0.10	0.20	ND	Passed
Mevinphos	0.10	1.00	ND	Passed
Metalaxyl	0.10	0.20	ND	Passed
Methiocarb	0.10	0.20	ND	Passed

DF New York One, Inc. - Jeeter

Order No.: ONYJTR0116-0011580
1050 Lower South St
New York, 10567
lpalazzo@jeeter.com
4424037621
OCM-PROC-24-000153-P2

Sample: SNYJTR0116-PEIP-0023026

Strain: Maui Wowie, Unit Weight: 2.0000g
Batch#: DFNY-MAUWOW-010926, Batch Size: 892 Units, LOT ID:
1A41203000001FC000000374
Date Sampled: 01/16/2026 16:44
Sample Received: 01/16/2026 17:52
Report Created: 01/29/2026 11:38
Sampling SOP 204 ♦

Infused Jeeter XL 2G Preroll Maui Wowie

Plant, Enhanced/Infused Preroll

Metrc Batch: 1A41203000001FC000000372, Metrc Sample: 1A41203000001FC000000374, Metrc Manifest: 0000061015



Analyte	LOQ (ug/g)	Limit (ug/g)	ug/g	Status
Methomyl	0.10	0.40	ND	Passed
MGK-264	0.01	0.20	ND	Passed
Myclobutanil	0.10	0.20	ND	Passed
Naled	0.10	0.50	ND	Passed
Oxamyl	0.10	1.00	ND	Passed
Paclobutrazol	0.10	0.40	ND	Passed
Permethrin	0.10	0.20	ND	Passed
Phosmet	0.10	0.20	ND	Passed
Piperonyl Butoxide	0.10	2.00	ND	Passed
Prallethrin	0.10	0.20	ND	Passed
Propiconazole	0.10	0.40	ND	Passed
Propoxur	0.10	0.20	ND	Passed
Pyrethrins	0.07	1.00	ND	Passed
Pyridaben	0.10	0.20	ND	Passed
Spinetoram	0.10	1.00	ND	Passed
Spinosyn AD	0.10	0.20	ND	Passed
Spiromesifen	0.10	0.20	ND	Passed
Spirotetramat	0.10	0.20	ND	Passed
Spiroxamine	0.10	0.20	ND	Passed
Tebuconazole	0.10	0.40	ND	Passed
Thiacloprid	0.10	0.20	ND	Passed
Thiamethoxam	0.10	0.20	ND	Passed
Trifloxystrobin	0.10	0.20	ND	Passed
Captan		1.00	TIC	Passed
Methyl Parathion	0.10	0.20	ND	Passed
Chlordane	0.10	1.00	ND	Passed
Chlorfenapyr	0.10	1.00	ND	Passed
PCNB	0.10	1.00	ND	Passed
Azadirachtin	0.12	1.00	ND	Passed
Chlormequat Chloride	0.02	1.00	ND	Passed

Report notes:

LOQ = Limit of Quantitation. Unless otherwise stated all quality control tests performed within specifications established by the Laboratory. If captan, chlormequat chloride, or MGK-264 are reported, they are tentatively identified, but not quantitatively confirmed. ND = Not Detected; NT = Not Tested; NR = Not Reported. "TIC" means tentatively identified, but not quantitatively confirmed.